



ADMINISTRATION OF ENTERAL FEEDING USING KANGAROO JOEY PUMP IN CONTINUOUS MODE

IMPORTANT:

- CHANGE FEEDING BAG EVERY 24 HRS.
- ONLY 8 HOURS OF FORMULA SHOULD BE POURED INTO BAG AT A TIME.
- CARRYING CASE IS MACHINE WASHABLE. USE GENTLE CYCLE.
- A FULLY CHARGED PUMP WILL RUN APPROXIMATELY 18 HOURS @ 125ML/HR

SUPPLIES:

- Formula
- Kangaroo Joey pump
- Kangaroo Joey feeding set (feeding bag)
- Backpack (if applicable) or IV pole
- 60 cc. syringe for flushing

PROCEDURE:

1. Gather supplies and equipment. Clean work area.
2. Perform hand hygiene.
3. Place Kangaroo Joey pump on a stable surface or attach it to the IV pole using the pole clamp.
NOTE: If using IV pole, it is important to secure the pump 6 inches or less from the feeding bag.
4. Remove Kangaroo feeding bag with tubing from package.
5. Fill the bag with the proper amount of formula and close bag securely at the top.
6. Turn pump on. The pump screen will display **PRIOR SETTINGS**. For your first home dose, press ►**CLEAR SETTINGS** key. After initial home dose, you will press ►**KEEP SETTINGS** key. The pump screen will display **LOAD A SET**.
7. Open the blue transparent door on the top of the pump.
8. Load the pump set into the pump by grasping finger tab on the set valve and inserting it into pump valve (as illustrated on top of pump). Grasp the black ring retainer and stretch the tubing counterclockwise around the pump rotor. Lower the black retainer into the slot (as illustrated on top of pump). Close the blue transparent door when the tubing is in place. The pump screen will display **SET LOADED**.



9. Press ► **PRIME** on the menu bar. Then press ► **AUTO PRIME**. The pump will automatically prime the tubing. When tubing is fully primed, the screen will display **AUTO PRIME COMPLETE**. You may press and hold ► **PRIME** key until tubing is fully primed. Then press ► **DONE**. To discontinue priming, press ► **STOP** to cancel **AUTO PRIME**.
10. The Kangaroo Joey pump will be preset in **CONTINUOUS MODE**. In continuous mode, using a “Feed Only Set”, settings that can be adjusted include **FEED RATE** and **VTBD** (Volume to be delivered). To change the rate (ml/hour), press ► **ADJUST FEED** key then press the ► **FEED RATE** key to set the correct rate. (Rates range from 1ml/hr – 400ml/hr). Press ► **ENTER** key to save the prescribed rate.
11. Next press ► **FEED VTBD** key and set the volume to be delivered. When prescribed volume is entered, press ► **ENTER** key to save prescribed volume.
12. Flush your feeding tube as instructed by your nurse.
13. Connect the tubing to your feeding tube by inserting tip of the tubing into the end of your feeding tube attachment. You may secure the connection with a piece of tape if necessary.
14. Press ► **RUN** key to start the pump. The pump screen will display **RUNNING**, and the droplet icon will move vertically on the screen. The feed rate will also display. The green LED indicator will be illuminated.
15. If the pump is for ambulatory use, place it and the feeding bag into the back pack.
16. The feeding can be interrupted by pressing ► **HOLD key**, which will pause the pump. The pump screen will display **HOLDING** and the yellow LED indicator will be illuminated. The pump will alarm if in **HOLDING** pause mode for 10 minutes or more. Press ► **RUN** to restart the pump.
17. When the formula has infused and feeding is complete, the pump will alarm. Press and hold ► **POWER DOWN** key to turn pump off.
18. Perform hand hygiene.
19. Disconnect the tubing from your feeding tube.
20. Flush your feeding tube as instructed by your nurse.
21. Plug the Kangaroo Joey enteral feeding pump into an A/C power outlet for pole mounted operation. Recharge the unit daily after using it as an ambulatory pump.



ENTERAL PRODUCT STORAGE GUIDELINES:

General Temperature Guidelines

General recommended storage temperature ranges for your nutritional product is between 32 and 95 degrees Fahrenheit (F). The most desirable temperature range for storage of unopened containers is room temperature, between 55 and 75 degrees F. Storage at these temperatures will assure the highest-quality product.

Unopened products should be stored in a cool, dry area. Prolonged exposure to temperatures below 32 degrees F or to direct heat above 95 degrees F could affect the physical consistency of the product.

Extreme Storage Conditions

Do not freeze formula products or store at greater than 95 degrees F. Excess temperatures, even for short time periods, can cause physical changes in the products, rendering them undesirable or unusable.

Additional Powder Storage Recommendations

Powdered formula should be used within one month from opening. Although the powders will not "spoil" from a microbiological perspective, some of the vitamins and fatty acids will degrade over time when exposed to oxygen. Powdered nutritional products should be stored in a cool, dry area. Opened powder cans should be covered; they should not be stored in the refrigerator.