



1001 Oakdale Road, Oakdale, PA 15071

PRODUCT RECALL

Dear Valued Patient:

Effective immediately, BD Medical has issued a voluntary recall of their **SafeStep Huber needle sets. SafeStep Huber needles in your home will need to be checked for specific lot numbers being recalled.**

Why are we doing this?

Product Recall. BD Medical recently conducted investigations in response to a concern with the extension tubing potentially developing cracks or breaks. The concern is minimal and unlikely, yet BD Medical is taking this precaution in the interest of public health.

For additional information, please view the Chartwell inventory being recalled on the next several pages. You can also find more information about the recall on BD's official Recall Notification document here: <https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>.

When will this happen?

- Effective immediately

What do I need to do?

- ❖ Contact Chartwell at **1-800-755-4704**
 - **choose option 3**
 - **then option 2**
 - then enter your **team number**
- ❖ We will review SafeStep lot numbers with you to determine if they should be disposed of or are safe to use.
- ❖ Chartwell will replenish your SafeStep Huber Needle stock in your next delivery.
- ❖ Your home nurse may also review your stock to ensure all recalled needles are disposed.

Thank you for your prompt attention and cooperation with this matter.

Chartwell Pharmacy Team

HUBER NDL 22G X 3/4" (LH0029) SAFESTEP**HCN: 3-33-691****MFG Catalog: LH-0029**

Lot	Expiration
ASCYF001	10/31/2021
ASCYF017	11/3/2021
ASDRF018	4/6/2022
ASDRF031	4/9/2022
ASDRF040	4/13/2022
ASDSF065	5/27/2022
ASDSF070	5/29/2022
ASDTF004	6/1/2022
ASDTF007	6/4/2022
ASDVF048	8/19/2022
ASDVF054	8/21/2022
ASDXF058	10/14/2022
ASDXF060	10/15/2022
ASDXF066	10/17/2022
ASDYF071	11/20/2022
ASDYF077	11/23/2022
ASDYF098	11/29/2022
ASDZF010	12/3/2022
ASDZF017	12/6/2022
ASDZF024	12/9/2022
ASDZF035	12/12/2022
ASDZF036	12/13/2022
ASDZF046	12/17/2022
ASDZF071	12/24/2022
ASDZF081	12/26/2022
ASDZF093	12/30/2022
ASENF002	1/2/2023
ASERF025	4/5/2023
ASERF032	4/7/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 22G X 1" (LH0030) SAFESTEP**HCN: 3-33-689****MFG Catalog: LH-0030**

Lot	Expiration
ASCYF015	11/2/2021
ASCYF016	11/2/2021
ASCZF027	12/14/2021
ASDRF043	4/15/2022
ASDRF044	4/15/2022
ASDRF054	4/16/2022
ASDRF055	4/17/2022
ASDRF056	4/17/2022
ASDVF058	8/23/2022
ASDVF063	8/25/2022
ASDXF020	10/9/2022
ASDXF045	10/12/2022
ASENF011	1/5/2023
ASENF014	1/6/2023
ASERF004	4/1/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 20G X 3/4" (LH0031) SAFESTEP**HCN: 3-33-688****MFG Catalog: LH-0031**

Lot	Expiration
ASCYF089	11/24/2021
ASCYF090	11/25/2021
ASCYF091	11/26/2021
ASCYF092	11/26/2021
ASCZF009	12/5/2021
ASCZF010	12/5/2021
ASCZF011	12/6/2021
ASCZF022	12/12/2021
ASCZF023	12/13/2021
ASCZF039	12/18/2021
ASCZF040	12/19/2021
ASCZF045	12/22/2021
ASCZF054	12/26/2021
ASDPF016	2/19/2022
ASDQF035	3/14/2022
ASDQF054	3/22/2022
ASDQF055	3/24/2022
ASDQF060	3/27/2022
ASDQF064	3/29/2022
ASDRF006	4/1/2022
ASDRF014	4/3/2022
ASDRF019	4/7/2022
ASDRF032	4/10/2022
ASDRF039	4/13/2022
ASDRF069	4/19/2022
ASDRF073	4/22/2022
ASDRF087	4/25/2022
ASDRF098	4/28/2022
ASDSF003	5/3/2022
ASDSF010	5/6/2022
ASDSF016	5/9/2022
ASDSF019	5/10/2022
ASDSF022	5/12/2022
ASDSF033	5/16/2022
Continued on page 4	

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 20G X 3/4" (LH0031) SAFESTEP CONTINUED

HCN: 3-33-688

MFG Catalog: LH-0031

Lot	Expiration
ASDSF047	5/18/2022
ASDSF050	5/19/2022
ASDSF057	5/23/2022
ASDSF061	5/26/2022
ASDSF069	5/29/2022
ASDTF005	6/1/2022
ASDTF008	6/4/2022
ASDTF022	6/9/2022
ASDVF051	8/20/2022
ASDVF060	8/29/2022
ASDVF071	8/26/2022
ASDVF079	8/28/2022
ASDVF091	8/29/2022
ASDWF004	9/1/2022
ASDWF013	9/4/2022
ASDWF018	9/6/2022
ASDWF034	9/10/2022
ASDWF035	9/10/2022
ASDWF036	9/10/2022
ASDWF040	9/10/2022
ASDWF041	9/10/2022
ASDWF042	9/10/2022
ASDWF046	9/11/2022
ASDWF047	9/11/2022
ASDWF048	9/11/2022
ASDWF052	9/11/2022
ASDWF053	9/11/2022
ASDWF054	9/11/2022

Lot	Expiration
ASDWF058	9/11/2022
ASDWF059	9/11/2022
ASDWF060	9/11/2022
ASDWF080	9/16/2022
ASDWF090	9/20/2022
ASDWF099	9/21/2022
ASDWF121	9/27/2022
ASDWF126	9/29/2022
ASDXF006	10/5/2022
ASDXF014	10/8/2022
ASDXF053	10/13/2022
ASDXF063	10/16/2022
ASDXF070	10/19/2022
ASDXF078	10/22/2022
ASDXF092	10/24/2022
ASDXF096	10/25/2022
ASDXF101	10/31/2022
ASDXF103	10/28/2022
ASDXF106	10/29/2022
ASDYF001	10/31/2022
ASDYF004	11/1/2022
ASDYF006	11/3/2022
ASDYF009	11/4/2022
ASDYF010	11/4/2022
ASDYF015	11/6/2022
ASDYF019	11/8/2022
ASDYF023	11/9/2022
ASDYF029	11/10/2022

Lot	Expiration
ASDYF033	11/12/2022
ASDYF037	11/13/2022
ASDYF044	11/14/2022
ASDYF050	11/16/2022
ASDYF054	11/17/2022
ASDYF060	11/18/2022
ASDYF061	11/19/2022
ASDYF074	11/21/2022
ASDYF082	11/24/2022
ASDYF093	11/27/2022
ASDZF004	12/1/2022
ASDZF013	12/4/2022
ASDZF058	12/20/2022
ASDZF069	12/23/2022
ASEPF012	2/21/2023
ASEPF015	2/23/2023
ASEPF017	2/24/2023
ASEPF021	2/26/2023
ASEQF001	3/12/2023
ASEQF022	3/6/2023
ASEQF026	3/8/2023
ASEQF031	3/9/2023
ASEQF066	3/17/2023
ASEQF083	3/21/2023
ASEQF092	3/24/2023
ASEQF110	3/27/2023
ASEQF132	3/31/2023
ASERF013	4/2/2023
ASERF022	4/3/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 20G X 1" (LH0032) SAFESTEP**HCN: 3-33-567****MFG Catalog: LH-0032**

Lot	Expiration
ASDNF034	1/19/2022
ASDNF049	1/23/2022
ASDNF055	1/27/2022
ASDPF015	2/18/2022
ASDPF017	2/19/2022
ASDPF021	2/21/2022
ASDPF024	2/23/2022
ASDPF025	2/23/2022
ASDPF026	2/23/2022
ASDQF017	3/5/2022
ASDRF066	4/18/2022
ASDUF120	7/30/2022
ASDVF004	8/1/2022
ASDVF009	8/4/2022
ASDVF013	8/5/2022
ASDVF016	8/6/2022
ASDVF025	8/11/2022
ASDVF028	8/12/2022
ASDVF032	8/14/2022
ASDVF052	8/20/2022
ASDXF048	10/12/2022
ASDXF052	10/13/2022
ASDXF061	10/15/2022
ASDXF068	10/18/2022
ASDXF077	10/21/2022
ASDXF091	10/24/2022
ASDXF100	10/27/2022
ASDXF107	10/29/2022
ASDYF083	11/25/2022
ASDYF088	11/27/2022
ASEPF008	2/18/2023
ASEPF011	2/20/2023
ASEPF014	2/21/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 19G X 1" WITHOUT YINJECTION SITE

HCN: 3-33-566

MFG Catalog: LH-0034



Lot	Expiration
ASCYF011	11/2/2021
ASCYF012	11/2/2021
ASDRF099	4/29/2022
ASDRF100	4/29/2022
ASDRF101	4/29/2022
ASDSF001	5/2/2022
ASDSF005	5/3/2022
ASDXF010	10/7/2022
ASDXF011	10/7/2022
ASDXF016	10/8/2022
ASDXF079	10/22/2022
ASENF051	1/12/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 22G X 1/2" (LH0035) SAFESTEP**HCN: 3-33-660****MFG Catalog: LH-0035**

Lot	Expiration
ASCYF013	11/2/2021
ASCYF014	11/2/2021
ASCZF029	12/14/2021
ASDVF034	8/14/2022
ASDVF035	8/15/2022
ASDVF038	8/15/2022
ASDVF042	8/16/2022
ASDVF044	8/18/2022
ASDXF054	10/13/2022
ASDXF055	10/13/2022
ASENF030	1/8/2023
ASENF033	1/9/2023
ASENF034	1/9/2023
ASENF042	1/10/2023

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>

HUBER NDL 20G X 1.5" (LH0038) SAFESTEP

HCN: 3-33-679

MFG Catalog: LH-0038



Lot	Expiration
ASDPF043	2/27/2022
ASDPF044	2/27/2022
ASDPF045	2/27/2022
ASDPF046	2/27/2022
ASDQF001	2/28/2022
ASDVF077	8/27/2022
ASDVF078	8/28/2022
ASDXF044	10/11/2022
ASDXF047	10/12/2022

For more information, please visit:

<https://www.bd.com/en-us/support/recall-notifications/recall-notification-powerloc-max-and-safestep>